



Description	Kluge & Fielitz GmbH Rehabilitation Cleaning and Disinfection Systems
Product:	Bed washing system type KF900/KF 900T
 	
<b>Features</b>	
Suitable for	Hospital beds, wheelchairs, walkers, AD mattresses, etc.
Capacity	for example 1 hospital bed, 3 wheelchairs
Cleaning programs	Duration 8,10 or 12 min.
Disinfection programs	2 (Duration 15 minutes)
Custom	yes
Washing programs possible available with condensation drying	Serie
Washing temperature	Adjustable up to 60 degrees °C
mobile	yes
Water consumption	35-45 litres
Tap water	min. 2 bar
Supply or waste water connection	Faucet 1/2" with external thread 3/4" or drain connection 1/2"
Electrical connection conditions	Socket 400 Volt / 32 A or 63Ass
Energy consumption	2.8 kW per wash cycle
<b>Service</b>	
Touch panel with control display	yes
Viewing Work Information	yes
Technical displays selectable	yes
Number of washing cycles (touch panel display)	yes
USB interface for information	yes
Program display	yes
<b>Safety</b>	
Emergency stop switch in the (Hand Area)	yes
During operation, open door-automatic stop	yes
Canister detergent empty - shut-off / display	yes
Noise level in dB(A)	62
Disinfection test possible for customers	yes
<b>Dimensions</b>	
Dimensions LxWxH	5400 mm x 2110 mm x 2810 mm (with technical part right or left)
Dimensions LxWxH (in the washing chamber)	2450 mm x 1430 mm x 1950 mm
Dimensions selectable / adjust	yes
Delivery in modules / assembly at the customer's site	yes
put in a pit	yes
Entrance height	WA inserted into pit - 0 mm; 170 mm without wheels; 270 mm with wheels
Weight	2500kg (depending on equipment)
<b>Maintenance</b>	
Lifetime at least	8-10 years
Number of washing cycles / year	Up to 3,000
Maintenance information is displayed in advance	yes
Filter easy to clean - without tools	yes
Materials for filters	Stainless steel
Program changes - at the customer's request	yes
Error display	Yes (indicator light / honking sound)
<b>Other auxiliary devices</b>	
Trolley	yes
Transport trolleys according to customer requirements	yes
Validated fogging disinfection	for foam and sensitive materials - on customer request
<b>Standards</b>	
CE, Medical Device Class II a	93/42/EEC; MDR; ISO 15883; DIN 13485
MOT	yes
<b>Warranty</b>	
Years (extension possible)	1